



DECLARATION OF CONFORMITY

DoC#: TF-1612, Rev 03.1

Legal Manufacturer: Innovacon, Inc.
Legal Manufacturer's Address: 9975 Summers Ridge Road
San Diego, CA 92121 USA

Declares that the product
Product Name and Model(s)

DAM-101	SureStep™ AMP One Step Amphetamine Test Strip (Urine) SureStep™ AMP 1000 Urine Drug Test Strip
DAM-102	SureStep™ AMP One Step Amphetamine Test Device (Urine) SureStep™ AMP 1000 Urine Drug Test Cassette
DAM-A101	SureStep™ AMP 300 One Step Amphetamine Test Strip (Urine) SureStep™ AMP 300 Urine Drug Test Strip
DAM-A102	SureStep™ AMP 300 One Step Amphetamine Test Device (Urine) SureStep™ AMP 300 Urine Drug Test Cassette
DAM-U101	SureStep™ AMP 500 One Step Amphetamine Test Strip (Urine) SureStep™ AMP 500 Urine Drug Test Strip
DAM-U102	SureStep™ AMP 500 One Step Amphetamine Test Device (Urine) SureStep™ AMP 500 Urine Drug Test Cassette

as described above are in conformity with the requirements of the standards listed in
Appendix 1, Applicable Standards and Guidelines.

Additional Information:

EC Representative's Name: Medical Device Safety Service GmbH
EC Representative's Address: Schiffgraben 41
30175 Hanover, Germany

Manufacturing Site: ABON Biopharm (Hangzhou) Co., Ltd.
#198, 12th Street East
Hangzhou Economic and Technological Development Area
310018 Hangzhou, PR China

Management System: MAN-003, Quality System Manual,

Quality System Certificate No: SX 60149668 0001, valid until 2021-08-24

Conformity Pathway: Annex III

Classification: Article 9, Section 1, Other IVD

EDMA Code: 12.70.09.02 – Amphetamine/Methamphetamine Specific
(+Ecstasy) - RT & POC



Abbott

This Declaration of Conformity is issued under the sole responsibility of Innovacon, Inc. I, the undersigned, hereby declare on behalf of the manufacturer, Innovacon, Inc., that the medical devices specified above conform with the Directive 98/79/EC on in vitro diagnostic medical devices and pertinent essential requirements.

Crystal Lee
Specialist II, Regulatory Affairs

San Diego, California, USA
Date: 12JUN2020



Appendix 1 to DOC # TF-1612		
Applicable Standards and Guidelines		
Category	Name	Number: Date Issued
General	In Vitro Diagnostic Device Directive	98/79/EC: 27 Oct 1998
	Medical Devices – Quality management systems – Requirements for regulatory purposes	EN ISO 13485:2016
Risk	Medical Devices – Application of risk management to medical devices	EN ISO 14971:2019
	Elimination or reduction of risk of infection related to in vitro diagnostic medical devices	EN 13641:2002
Labeling	Symbols for use in the labeling of medical devices	EN 15223-1:2016
	In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) – Part 1: Terms, definitions and general requirements	EN ISO 18113-1:2011
	In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) – Part 2: In vitro diagnostic reagents for professional use	EN ISO 18113-2:2011
Performance	Performance evaluation of in vitro diagnostic medical devices	EN 13612:2002/AC:2002
	In vitro diagnostic medical devices – Evaluation of stability of in vitro diagnostic reagents	EN ISO 23640:2015