

**Konformitätserklärung
Declaration of Conformity**

Wir

We

**B. Braun Melsungen AG
Carl-Braun-Straße 1
34212 Melsungen
Deutschland/Germany**erklären in eigener Verantwortung,
dass das/die Produkt/e**Solofix® Safety Universal
Solofix® Safety Fine
Solofix® Safety Neonat**
Lanzette, Blut, steril

(Artikelnummern siehe Anlage I)

mit den Anforderungen der folgenden Richtlinie
übereinstimmt/übereinstimmenRichtlinie 93/42/EWG des Rates vom 14. Juni
1993 über Medizinprodukte,
geändert durch Richtlinie 2007/47/EG**Konformitätsbewertungsverfahren**
nach Anhang II
ohne Abschnitt 4
der oben genannten Richtlinie**Klassifizierung**
gemäß Anhang IX der oben genannten
Richtlinie:
Klasse IIa**Benannte Stelle**
TÜV SÜD Product Service GmbH
Ridlerstraße 65
80339 München
Deutschland
Kennnummer 0123**Datum der ersten CE-Kennzeichnung**
2008-05**Gültig bis**
2024-05-26hereby declare in our own responsibility
that the product/s**Solofix® Safety Universal
Solofix® Safety Fine
Solofix® Safety Neonat**
Lancets, blood, sterile

(article numbers see attachment I)

is/are in compliance with the following directive

Council Directive 93/42/EEC of 14 June 1993
concerning Medical Devices,
amended by Directive 2007/47/EG**Conformity assessment procedure**
according to annex II
without part 4
of the Directive named above**Classification**
according to annex IX of the Directive named
above:
Class IIa**Notified body**
TÜV SÜD Product Service GmbH
Ridlerstraße 65
80339 München
Germany
Identification number 0123**Date of first CE-marking**
2008-05**Valid until**
2024-05-26

Anlage I / Attachment I

Art.-Nr. / Art. No.	Produktname / Product Name	Klasse / Class
6183000	Solofix® Safety Universal	Ila
6183010	Solofix® Safety Fine	Ila
6183020	Solofix® Safety Neonat	Ila

Amendment Information

Version	Description of the changes
09	Yearly revision acc. to TÜV audit 05-2009 Version 1: BBMAG OPM; Version 2: BBMAG
10	Adaption to SOP HO-DDDD-M-5-2-05-715 Add art. no 6183040, 60183050 again (SAP status M4/V4 30.06.2014 and 17.09.2012, articles have not expired yet)
11	Adaption of conformity assessment procedure to “according to annex II without part 4”
12	Delete Out of Market art. no. 6183040, 6183050
13	Adaptation of validity to the MDR-deadline

Title: Declaration of Conformity - 084-005 - Solofix Safety Initiator: Anna Heil

This document is signed electronically in compliance with the B. Braun electronic signature policies and procedures by following persons:

UserName: Heil, Anna (heilaade)
Title: Administrator Regulatory Affairs
Date: Wednesday, 01 April 2020, 15:50 W. Europe Daylight Time
Meaning: Document signed as Author

UserName: Moeller, Thomas (moelthde)
Title: HC-QM-DE08A - Vice President QM Systems & Quality Compliance
Date: Thursday, 02 April 2020, 08:21 W. Europe Daylight Time
Meaning: Approve Document

UserName: Gaudin, Hans-Ulrich (gaudhade)
Title: OM-RADE08S Head of Regulatory Affairs OPM DE
Date: Tuesday, 07 April 2020, 11:15 W. Europe Daylight Time
Meaning: Approve Document
