



Vitalograph (Ireland) Limited, Gort Road Business Park, Ennis, Co. Clare, Ireland
Telephone: 065 686 4100 Fax: 065 682 9289

Declaration of Conformity

(To the requirements of 93/42 EEC 169, volume 36, as amended)


Vitalograph hereby ensures and declares that the following Products & their Accessories, bearing the CE Mark, meet the following provisions of the Directive:

- Class 2a devices meet the provisions of Annex II per Article 11, Sect 3a.
- Class 1 devices comply with Annex VII per Article 11 Sect 5. Devices classified as Class 1, with a measuring function, comply with Annex V per Annex VII Sect 5.

| MODEL N ^o | PRODUCT NAME | GMDN | CLASS |
|----------------------|-------------------------|--------|-------|
| 2020 | Mouthpiece | P44545 | 1** |
| 2024 | SafeTway | P44545 | 2a** |
| 2030 | Noseclip | P10907 | 1** |
| 2040 | Precision Syringe | P17250 | 1* |
| 2110 | PEF/FEV Diary | P46906 | 2a |
| 2120 | Hand Held | P13680 | 2a |
| 2130 | Base Station Spirometer | P13680 | 2a |
| 2150 | Gold Standard | P13680 | 2a |
| 2160 | Gold Standard + | P13680 | 2a |
| 2200 | ResusBag | P17591 | 2a |
| 2400 | Resuscitation Kit | P36690 | 2a |
| 2410 | Resuscitation Outfit | P36690 | 2a |
| 2510 | Emergency Aspirator | P47368 | 1 |
| 2820 | BVF | P44545 | 2a** |
| 4000 | Respiratory Monitor | P46906 | 2a |
| 4300 | Peak Flow Meter | P46872 | 1* |
| 4530 | Inhaler Compliance | | 1 |
| 4500 | AIM | P42629 | 1* |
| 6000 | Alpha | P13680 | 2a |
| 6300 | Micro | P46906 | 2a |
| 6600 | Compact | P13680 | 2a |
| 6800 | Pneumotrac | P13680 | 2a |
| 7000 | Spirotrac | P13680 | 2a |
| 7100 | VitaloJAK | P16305 | 2a |
| 8000 | Web Portal | | 1** |
| 8400 | Inhaler Trainer | | 1 |

* Indicates Class 1 with a measuring function.

** Indicates Accessory to one or more of these products


Tom J Healy

Regulatory Affairs / QA Manager.

Date: 15th January 2014